

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	0.000 °C	0.3 °C	0.300 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.047 °C	10.3 °C	0.253 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.205 °C	20.3 °C	0.095 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.002 °C	29.7 °C	-0.302 °C	± 0.5 °C	± 4.0E-01 °C