

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
0 °C	0.000 °C	0.4 °C	0.400 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.047 °C	10.3 °C	0.253 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.205 °C	20.3 °C	0.095 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.002 °C	29.7 °C	-0.302 °C	± 0.5 °C	± 4.0E-01 °C