

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.614 °C	30.6 °C	-0.014 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	29.0 %HR	-1.570 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.120 %HR	53.0 %HR	2.880 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.570 %HR	69.0 %HR	-1.570 %HR	± 5.0 %HR	± 1.9E+00 %HR