

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>	
1.0015 ml	1 ml	-0.002 ml	± 0.05 ml	±	7.8E-02 µl
5.0012 ml	5 ml	-0.001 ml	± 0.13 ml	±	3.0E-01 µl
10.0023 ml	10 ml	-0.002 ml	± 0.12 ml	±	5.9E-01 µl