

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.140 °C	0.0 °C	-0.140 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.189 °C	20.3 °C	0.111 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.096 °C	35.3 °C	0.204 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	29.1 %HR	-1.163 %HR	± 5.0 %HR	± 1.9E+00 %HR
55 %HR	55.233 %HR	54.4 %HR	-0.833 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	80.161 %HR	78.6 %HR	-1.561 %HR	± 5.0 %HR	± 1.9E+00 %HR