

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
23 °C	23.093 °C	23.1 °C	-0.033 °C	± 0.2 °C	± 4.0E-01 °C
25 %HR	25.458 %HR	25.6 %HR	0.142 %HR	± 2.0 %HR	± 1.9E+00 %HR
50 %HR	50.110 %HR	50.3 %HR	0.190 %HR	± 2.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	71.9 %HR	1.452 %HR	± 2.0 %HR	± 1.9E+00 %HR