
Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
600 RPM	600.0 RPM	0.0 RPM	± 0.54 RPM	±	5.8E-02 RPM
3000 RPM	3000.0 RPM	0.0 RPM	± 0.68 RPM	±	5.8E-02 RPM
9000 RPM	9000.1 RPM	0.1 RPM	± 1.0 RPM	±	7.0E-01 RPM
60000 RPM	60001 RPM	1.0 RPM	± 4.1 RPM	±	9.0E-01 RPM
90000 RPM	90002 RPM	2.0 RPM	± 5.9 RPM	±	9.3E-01 RPM