

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.123 °C	29.8 °C	-0.323 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.417 %HR	33.2 %HR	2.783 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.558 %HR	51.3 %HR	1.742 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.478 %HR	69.3 %HR	-1.178 %HR	± 3.0 %HR	± 1.9E+00 %HR