

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.123 °C	29.9 °C	-0.223 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.417 %HR	35.0 %HR	4.583 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.558 %HR	53.0 %HR	3.442 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.478 %HR	75.0 %HR	4.522 %HR	± 5.0 %HR	± 1.9E+00 %HR