
Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
100.009 ml	100 ml	-0.009 ml	± 0.10 ml	± 2.9E+00 µl
