

<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
2.5034 ml	2.5000 ml	-0.0034 ml	± 0.05 ml	± 1.1E-01 µl
12.5080 ml	12.5000 ml	-0.0080 ml	± 0.05 ml	± 3.9E-01 µl
25.0092 ml	25.0000 ml	-0.0092 ml	± 0.05 ml	± 7.8E-01 µl