

---

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
499.59700 ml	500 ml	0.40300 ml	± 0.25 ml	± 4.2E+01 µl

---