

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.450 °C	-25.3 °C	0.150 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	0.020 °C	0.2 °C	0.180 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.314 °C	5.2 °C	-0.114 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.428 °C	10.3 °C	-0.128 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.600 °C	25.3 °C	-0.300 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.614 °C	30.3 °C	-0.314 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.050 °C	39.6 °C	-0.450 °C	± 0.5 °C	± 4.0E-01 °C

