

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.450 °C	-25.2 °C	0.250 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	0.020 °C	0.4 °C	0.380 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.314 °C	5.8 °C	0.486 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.428 °C	10.8 °C	0.372 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.600 °C	26.1 °C	0.500 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.614 °C	31.1 °C	0.486 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.050 °C	40.3 °C	0.250 °C	± 0.5 °C	± 4.0E-01 °C

