

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C | -25.450 °C | -25.0 °C | 0.450 °C | ± 0.7 °C | ± 4.0E-01 °C |
| 0 °C | 0.020 °C | 0.4 °C | 0.380 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 5 °C | 5.314 °C | 5.7 °C | 0.386 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 10 °C | 10.428 °C | 10.8 °C | 0.372 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 25 °C | 25.600 °C | 25.7 °C | 0.100 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 30 °C | 30.614 °C | 30.7 °C | 0.086 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 40 °C | 40.050 °C | 40.0 °C | -0.050 °C | ± 0.5 °C | ± 4.0E-01 °C |

