

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C | -25.180 °C | -24.7 °C | 0.480 °C | ± 0.7 °C | ± 4.0E-01 °C |
| 0 °C | 0.140 °C | 0.6 °C | 0.460 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 5 °C | 5.184 °C | 5.6 °C | 0.416 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 10 °C | 10.288 °C | 10.7 °C | 0.412 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 25 °C | 25.570 °C | 26.0 °C | 0.430 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 30 °C | 30.173 °C | 30.5 °C | 0.327 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 40 °C | 39.909 °C | 40.2 °C | 0.291 °C | ± 0.5 °C | ± 4.0E-01 °C |

