

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.180 °C	-25.5 °C	-0.320 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	0.140 °C	-0.2 °C	-0.340 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.184 °C	4.9 °C	-0.284 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.288 °C	9.9 °C	-0.388 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.570 °C	25.2 °C	-0.370 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.173 °C	29.9 °C	-0.273 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.909 °C	39.6 °C	-0.309 °C	± 0.5 °C	± 4.0E-01 °C

