

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.180 °C	-24.9 °C	0.280 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	0.140 °C	0.2 °C	0.060 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.184 °C	5.2 °C	0.016 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.288 °C	10.4 °C	0.112 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.570 °C	25.4 °C	-0.170 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.173 °C	30.3 °C	0.127 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.909 °C	39.6 °C	-0.309 °C	± 0.5 °C	± 4.0E-01 °C

