

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.180 °C	-24.7 °C	0.480 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	0.140 °C	0.4 °C	0.260 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.184 °C	5.5 °C	0.316 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.288 °C	10.6 °C	0.312 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.570 °C	25.9 °C	0.330 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.173 °C	30.5 °C	0.327 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.909 °C	40.2 °C	0.291 °C	± 0.5 °C	± 4.0E-01 °C

