

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.180 °C	-25.0 °C	0.180 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	0.140 °C	0.3 °C	0.160 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.184 °C	5.4 °C	0.216 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.288 °C	10.5 °C	0.212 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.570 °C	25.8 °C	0.230 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.173 °C	30.3 °C	0.127 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.909 °C	39.7 °C	-0.209 °C	± 0.5 °C	± 4.0E-01 °C

