

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.180 °C	-24.6 °C	0.580 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	0.140 °C	0.5 °C	0.360 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.184 °C	5.6 °C	0.416 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.288 °C	10.6 °C	0.312 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.570 °C	26.0 °C	0.430 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.173 °C	30.6 °C	0.427 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.909 °C	40.3 °C	0.391 °C	± 0.5 °C	± 4.0E-01 °C

