

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.240 °C	-25.1 °C	0.140 °C	± 0.95 °C	± 4.0E-01 °C
0 °C	-0.041 °C	-0.1 °C	-0.059 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	4.973 °C	5.1 °C	0.127 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.158 °C	10.2 °C	0.042 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	25.069 °C	25.0 °C	-0.069 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	30.072 °C	30.2 °C	0.128 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	40.301 °C	40.4 °C	0.099 °C	± 0.45 °C	± 4.0E-01 °C

