

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.240 °C	-24.6 °C	0.640 °C	± 0.95 °C	± 4.0E-01 °C
0 °C	-0.041 °C	0.2 °C	0.241 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	4.973 °C	5.0 °C	0.027 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.158 °C	10.1 °C	-0.058 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	25.069 °C	24.9 °C	-0.169 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	30.072 °C	29.8 °C	-0.272 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	40.301 °C	40.0 °C	-0.301 °C	± 0.45 °C	± 4.0E-01 °C

