

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.09975	0.250	± 0.6	± 4.3E-01 mm					
0 a 0.5	0.49973	0.270	± 0.8	± 4.3E-01 mm					
0 a 1	0.99971	0.290	± 1	± 4.3E-01 mm					
0 a 2	1.99965	0.350	± 1.4	± 4.7E-01 mm					
0 a 3	2.99944	0.560	± 1.8	± 4.7E-01 mm					
0 a 4	3.99934	0.660	± 2.2	± 4.7E-01 mm					
0 a 5	4.99925	0.750	± 2.6	± 4.7E-01 mm					
0 a 6	5.99899	1.010	± 3	± 3.4E-01 mm					
0 a 7	6.99879	1.210	± 3.4	± 3.4E-01 mm					
0 a 8	7.99858	1.420	± 3.8	± 3.4E-01 mm					
