

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
0 °C	0.240 °C	0.7 °C	0.460 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.165 °C	20.6 °C	0.435 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	34.615 °C	34.8 °C	0.185 °C	± 0.5 °C	± 4.0E-01 °C
# 10 %HR	10.774 %HR	12.2 %HR	1.426 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.701 %HR	48.4 %HR	-1.301 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	81.409 %HR	78.9 %HR	-2.509 %HR	± 3.0 %HR	± 1.9E+00 %HR