

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
0 °C	0.240 °C	0.4 °C	0.160 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.165 °C	20.5 °C	0.335 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	34.615 °C	35.1 °C	0.485 °C	± 0.5 °C	± 4.0E-01 °C
# 10 %HR	10.774 %HR	11.3 %HR	0.526 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.701 %HR	47.9 %HR	-1.801 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	81.409 %HR	78.8 %HR	-2.609 %HR	± 3.0 %HR	± 1.9E+00 %HR