

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.240 °C	0.5 °C	0.260 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.165 °C	20.6 °C	0.435 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	34.615 °C	35.1 °C	0.485 °C	± 0.5 °C	± 4.0E-01 °C
# 10 %HR	10.774 %HR	13.6 %HR	2.826 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.701 %HR	49.6 %HR	-0.101 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	81.409 %HR	78.5 %HR	-2.909 %HR	± 3.0 %HR	± 1.9E+00 %HR