

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 0 °C | 0.240 °C | 0.4 °C | 0.160 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 20 °C | 20.165 °C | 20.5 °C | 0.335 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 35 °C | 34.615 °C | 35.0 °C | 0.385 °C | ± 0.5 °C | ± 4.0E-01 °C |
| # 10 %HR | 10.774 %HR | 13.0 %HR | 2.226 %HR | ± 3.0 %HR | ± 1.9E+00 %HR |
| 50 %HR | 49.701 %HR | 48.5 %HR | -1.201 %HR | ± 3.0 %HR | ± 1.9E+00 %HR |
| 80 %HR | 81.409 %HR | 78.5 %HR | -2.909 %HR | ± 3.0 %HR | ± 1.9E+00 %HR |