

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.323 °C	29.8 °C	-0.523 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.813 %HR	34 %HR	4.187 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.693 %HR	54 %HR	3.307 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.407 %HR	79 %HR	3.593 %HR	± 5.0 %HR	± 1.9E+00 %HR