

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 30 °C | 30.323 °C | 30.1 °C | -0.223 °C | ± 1.0 °C | ± 4.0E-01 °C |
| 30 %HR | 29.813 %HR | 33 %HR | 3.187 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |
| 50 %HR | 50.693 %HR | 49 %HR | -1.693 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |
| 75 %HR | 75.407 %HR | 73 %HR | -2.407 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |