

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	1.935 °C	2.5 °C	0.565 °C	± 1.0 °C	± 4.0E-01 °C
8 °C	8.032 °C	8.6 °C	0.568 °C	± 1.0 °C	± 4.0E-01 °C
50 %HR	50.693 %HR	52 %HR	1.307 %HR	± 5.0 %HR	± 1.9E+00 %HR