

Certificado No. (Certificate #):

67857

Equipo (Instrument): Reloj Palpador

Fecha de Calibración (Calibration Date):

2026-04-02

Marca (Brand): Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0.0010 in	0.0010 in	0.0000 in	± 0.0005 in	± 2.9E-04 in
0.0050 in	0.0050 in	0.0000 in	± 0.0005 in	± 2.9E-04 in
0.0060 in	0.0060 in	0.0000 in	± 0.0005 in	± 2.9E-04 in
0.0070 in	0.0070 in	0.0000 in	± 0.0005 in	± 2.9E-04 in
0.0080 in	0.0080 in	0.0000 in	± 0.0005 in	± 2.9E-04 in
0.0090 in	0.0090 in	0.0000 in	± 0.0005 in	± 2.9E-04 in
0.0100 in	0.0100 in	0.0000 in	± 0.0005 in	± 2.9E-04 in
0.0200 in	0.0200 in	0.0000 in	± 0.0005 in	± 2.9E-04 in
0.0300 in	0.0300 in	0.0000 in	± 0.0005 in	± 2.9E-04 in