

Certificado No. (Certificate #):

67858

Equipo (Instrument): Reloj Palpador

Fecha de Calibración (Calibration Date):

2026-04-02

Marca (Brand): Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0.05 mm	0.05 mm	0 µm	± 8 µm	± 5.8E+00 µm
0.10 mm	0.10 mm	0 µm	± 8 µm	± 5.8E+00 µm
0.15 mm	0.15 mm	0 µm	± 8 µm	± 5.8E+00 µm
0.20 mm	0.20 mm	0 µm	± 8 µm	± 5.8E+00 µm
0.25 mm	0.25 mm	0 µm	± 8 µm	± 5.8E+00 µm
0.30 mm	0.30 mm	0 µm	± 8 µm	± 5.8E+00 µm
0.35 mm	0.35 mm	0 µm	± 8 µm	± 5.8E+00 µm
0.50 mm	0.50 mm	0 µm	± 8 µm	± 5.8E+00 µm
0.75 mm	0.75 mm	0 µm	± 8 µm	± 5.8E+00 µm