

Certificado No. (Certificate #): 67859

Equipo (Instrument): Micrómetro de Exterior

Fecha de Calibración (Calibration Date): 2026-04-02

Marca (Brand): Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
0.05000 in	0.05000 in	0.00000 in	± 0.0001 in	± 2.9E-05 in
0.10000 in	0.10000 in	0.00000 in	± 0.0001 in	± 2.9E-05 in
0.10030 in	0.10030 in	0.00000 in	± 0.0001 in	± 2.9E-05 in
0.10050 in	0.10050 in	0.00000 in	± 0.0001 in	± 2.9E-05 in
0.10300 in	0.10300 in	0.00000 in	± 0.0001 in	± 2.9E-05 in
0.10500 in	0.10500 in	0.00000 in	± 0.0001 in	± 2.9E-05 in
0.15000 in	0.15000 in	0.00000 in	± 0.0001 in	± 2.9E-05 in
0.20000 in	0.20000 in	0.00000 in	± 0.0001 in	± 2.9E-05 in
0.30000 in	0.30000 in	0.00000 in	± 0.0001 in	± 2.9E-05 in
0.50000 in	0.50000 in	0.00000 in	± 0.0001 in	± 2.9E-05 in
1.00000 in	1.00000 in	0.00000 in	± 0.0001 in	± 2.9E-05 in