

Certificado No. (Certificate #): 67863

Equipo (Instrument): Micrómetro de Interior

Fecha de Calibración (Calibration Date): 2026-04-02

Marca (Brand): Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
0.8000 in	0.7998 in	-0.0002 in	± 0.0003 in	± 1.2E-04 in
0.8500 in	0.8502 in	0.0002 in	± 0.0003 in	± 1.2E-04 in
0.9000 in	0.9002 in	0.0001 in	± 0.0003 in	± 1.2E-04 in
0.9500 in	0.9500 in	0.0000 in	± 0.0003 in	± 1.2E-04 in
1.0000 in	1.0002 in	0.0002 in	± 0.0003 in	± 1.2E-04 in