

Certificado No. (Certificate #): 67865

Equipo (Instrument): Micrómetro de Interior

Fecha de Calibración (Calibration Date): 2026-04-02

Marca (Brand): Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
1.2000 in	1.2000 in	0.0000 in	± 0.0003 in	± 1.2E-04 in
1.2500 in	1.2498 in	-0.0002 in	± 0.0003 in	± 1.2E-04 in
1.3000 in	1.3000 in	0.0000 in	± 0.0003 in	± 1.2E-04 in
1.4000 in	1.3998 in	-0.0002 in	± 0.0003 in	± 1.2E-04 in
1.4500 in	1.4498 in	-0.0002 in	± 0.0003 in	± 1.2E-04 in