

Certificado No. (Certificate #): 67866

Equipo (Instrument): Micrómetro de Interior

Fecha de Calibración (Calibration Date): 2026-04-02

Marca (Brand): Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
1.6000 in	1.6002 in	0.0002 in	± 0.0003 in	± 1.2E-04 in
1.6500 in	1.6498 in	-0.0002 in	± 0.0003 in	± 1.2E-04 in
1.7000 in	1.6998 in	-0.0002 in	± 0.0003 in	± 1.2E-04 in
1.8000 in	1.8002 in	0.0002 in	± 0.0003 in	± 1.2E-04 in
1.8500 in	1.8498 in	-0.0002 in	± 0.0003 in	± 1.2E-04 in