

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
49.90 mmHg	50 mmHg	0.10 mmHg	± 5 mmHg	±	#REF! mmHg
100.00 mmHg	100 mmHg	0.00 mmHg	± 5 mmHg	±	#REF! mmHg
150.00 mmHg	150 mmHg	0.00 mmHg	± 5 mmHg	±	#REF! mmHg
199.90 mmHg	200 mmHg	0.10 mmHg	± 5 mmHg	±	#REF! mmHg
249.80 mmHg	250 mmHg	0.20 mmHg	± 5 mmHg	±	#REF! mmHg
299.80 mmHg	300 mmHg	0.20 mmHg	± 5 mmHg	±	#REF! mmHg