

Certificado No. (Certificate #):

67875

Equipo (Instrument): Calibrador Pie de Rey

Fecha de Calibración (Calibration Date):

2026-04-01

Marca (Brand):

Mitutoyo

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Exterior / (Exterior)	0.5000 in	0.5000 in	0.0000 in	± 0.0012 in	± 3.5E-04 in
Exterior / (Exterior)	1.0000 in	1.0000 in	0.0000 in	± 0.0012 in	± 3.5E-04 in
Exterior / (Exterior)	3.0000 in	3.0000 in	0.0000 in	± 0.0012 in	± 3.5E-04 in
Exterior / (Exterior)	5.0000 in	5.0000 in	0.0000 in	± 0.0012 in	± 3.5E-04 in
Exterior / (Exterior)	7.5000 in	7.5000 in	0.0000 in	± 0.0012 in	± 3.5E-04 in
Interior / (Interior)	1.5000 in	1.5000 in	0.0000 in	± 0.0012 in	± 3.7E-04 in
Interior / (Interior)	5.0000 in	5.0000 in	0.0000 in	± 0.0012 in	± 3.7E-04 in
Profundidad / (Depth)	2.0000 in	1.9998 in	-0.0002 in	± 0.0012 in	± 7.4E-04 in
Profundidad / (Depth)	4.0000 in	3.9998 in	-0.0002 in	± 0.0012 in	± 4.7E-04 in