

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
52.65 μm	52.92 μm	0.27 μm	± 1.6 μm	± 5.8E-02 μm
155.23 μm	155.0 μm	-0.23 μm	± 4.7 μm	± 6.2E-02 μm
257.12 μm	257.6 μm	0.48 μm	± 7.7 μm	± 7.0E-02 μm
511.72 μm	511.6 μm	-0.12 μm	± 15.4 μm	± 9.7E-02 μm