

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1.005 mm	1.010 mm	5 μm	$\pm 10.05 \mu\text{m}$	\pm	5.8E+01 μm
1.5 mm	1.501 mm	1 μm	$\pm 15 \mu\text{m}$	\pm	5.8E+01 μm
2 mm	2.004 mm	4 μm	$\pm 20 \mu\text{m}$	\pm	5.8E+01 μm
4 mm	4.003 mm	3 μm	$\pm 40 \mu\text{m}$	\pm	5.8E+01 μm
5 mm	5.008 mm	8 μm	$\pm 50 \mu\text{m}$	\pm	5.8E+01 μm