

Patrón (Standard)		UBP (UIT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
7 Hz	25.23 mm/s-rms	25.3 mm/s-rms	0.07 mm/s-rms	± 0.58 mm/s-rms	± 5.9E-01 mm/s-rms
50 Hz	22.09 mm/s-rms	22.1 mm/s-rms	0.01 mm/s-rms	± 0.53 mm/s-rms	± 4.8E-01 mm/s-rms
100 Hz	11.03 mm/s-rms	11.1 mm/s-rms	0.07 mm/s-rms	± 0.37 mm/s-rms	± 3.4E-01 mm/s-rms
100 Hz	11.10 mm/s-rms	11.2 mm/s-rms	0.10 mm/s-rms	± 0.37 mm/s-rms	± 3.4E-01 mm/s-rms