

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.323 °C	30.5 °C	0.177 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.813 %HR	32.4 %HR	2.587 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.693 %HR	52.4 %HR	1.707 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.407 %HR	73.6 %HR	-1.807 %HR	± 3.0 %HR	± 1.9E+00 %HR