

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.323 °C	30.3 °C	-0.023 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.813 %HR	32.5 %HR	2.687 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.693 %HR	52.3 %HR	1.607 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.407 %HR	73.4 %HR	-2.007 %HR	± 3.0 %HR	± 1.9E+00 %HR