

| <b>Nominal</b> ( <i>Nominal</i> ) | <b>Patrón</b> ( <i>Standard</i> ) | <b>UBP</b> ( <i>UUT</i> ) | <b>Error</b> ( <i>Error</i> ) | <b>Exactitud</b> ( <i>Accuracy</i> ) | <b>Incertidumbre</b> ( <i>Uncertainty</i> ) |
|-----------------------------------|-----------------------------------|---------------------------|-------------------------------|--------------------------------------|---|
| 30 °C                             | 30.323 °C                         | 30.9 °C                   | 0.577 °C                      | ± 1.0 °C                             | ± 4.0E-01 °C                                |
| 30 %HR                            | 29.813 %HR                        | 34 %HR                    | 3.887 %HR                     | ± 5.0 %HR                            | ± 1.9E+00 %HR                               |
| 50 %HR                            | 50.693 %HR                        | 50 %HR                    | -1.193 %HR                    | ± 5.0 %HR                            | ± 1.9E+00 %HR                               |
| 75 %HR                            | 75.407 %HR                        | 73 %HR                    | -2.507 %HR                    | ± 5.0 %HR                            | ± 1.9E+00 %HR                               |