

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
15 °C	14.931 °C	14.6 °C	-0.331 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	19.914 °C	19.6 °C	-0.314 °C	± 1.0 °C	± 4.0E-01 °C
23 °C	23.023 °C	22.7 °C	-0.323 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	30.323 °C	30.1 °C	-0.223 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.813 %HR	32.0 %HR	2.187 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.693 %HR	48.0 %HR	-2.693 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.681 %HR	68.0 %HR	-1.681 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.407 %HR	75.0 %HR	-0.407 %HR	± 5.0 %HR	± 1.9E+00 %HR