
Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (M.P.E)	Incertidumbre (Uncertainty)	Regla de decisión
2.0006 ml	2 ml	-0.0006 ml	± 0.1.0 ml	± 3.5E-01 µl	Cumple
4.9908 ml	5 ml	0.0092 ml	± 0.1.0 ml	± 3.5E-01 µl	Cumple
9.9955 ml	10 ml	0.0045 ml	± 0.1.0 ml	± 3.5E-01 µl	Cumple