

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (M.P.E)	Incertidumbre (Uncertainty)	Regla de decisión
9.9999 ml	10 ml	0.0001 ml	± 0.5.0 ml	± 3.1E-01 µl	Cumple
25.0178 ml	25 ml	-0.0178 ml	± 0.5.0 ml	± 7.8E-01 µl	Cumple
50.0789 ml	50 ml	-0.0789 ml	± 0.5.0 ml	± 1.5E+00 µl	Cumple