

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (M.P.E)	Incertidumbre (Uncertainty)	Regla de decisión
10.0022 ml	10 ml	-0.0022 ml	± 0.5.0 ml	± 3.5E-01 µl	Cumple
50.0888 ml	50 ml	-0.0888 ml	± 0.5.0 ml	± 3.5E-01 µl	Cumple
99.9890 ml	100 ml	0.0110 ml	± 0.5.0 ml	± 3.5E-01 µl	Cumple