

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>E.M.P (M.P.E)</b>	<b>Incertidumbre (Uncertainty)</b>	<b>Regla de decisión</b>
10.0022 ml	10 ml	-0.0022 ml	± 0.50 ml	± 3.5E-01 µl	Cumple
50.0888 ml	50 ml	-0.0888 ml	± 0.50 ml	± 3.5E-01 µl	Cumple
99.9890 ml	100 ml	0.0110 ml	± 0.50 ml	± 3.5E-01 µl	Cumple